





embecta Compatibility Confirmation for Pen Needles and Pen Injector Manufacturer: Sanofi

Doc #:149OTH-0004-31  
Rev: D  
Doc Type: ZSP

and are additionally compatible with embecta AutoShield Duo™ and embecta SafeAssist™ in the following sizes:

- 0.30 mm (30G) x 5 mm
- 0.30 mm (30G) x 8 mm

This confirmation is based on the Technical Agreement previously signed between embecta (previously known as BD Diabetes Care) and SANOFI. Confirmation is supported by verification test results performed by embecta to meet the requirements of ISO 11608-2:2022, Section 9. This compatibility confirmation is updated as necessary.

<p>Anne Hillbrecht Lead Device External Manufacturing, Sanofi Mfg. Technologies LCM and Supply</p> <p>Questions regarding statements herein: Anne Hillbrecht Sanofi Lead Device External Manufacturing Anne.Hillbrecht@sanofi.com</p>	<p><b>Signature:</b> <i>Anne Hillbrecht</i></p> <p><b>Email:</b> anne.hillbrecht@sanofi.com</p> <p><small>Electronically signed by: Anne Hillbrecht Reason: I approve this document. Date: Oct 30, 2025 10:39:34 GMT+1</small></p>
---	--

<p>Cal McCarthy Director, Quality Operations, embecta</p> <p>Questions regarding statements herein: Cal McCarthy embecta Director, Quality Operation Cal.McCarthy@embecta.com</p>	<p><b>Signature:</b> <i>Cal McCarthy</i></p> <p><b>Email:</b> cal.mccarthy@embecta.com</p> <p><small>Electronically signed by: Cal McCarthy Reason: I approve this document. Date: Oct 28, 2025 08:52:49 GMT</small></p>
---	--

CONFIDENTIAL - THE SUBJECT MATTER CONTAINED IN THIS DOCUMENT IS SOLELY FOR THE USE OF embecta  
The reader is responsible for ensuring that they use only the current revision of this document for compliance purposes



embecta Compatibility Confirmation for Pen Needles and Pen  
Injector Manufacturer: Sanofi

Doc #:149OTH-0004-31  
Rev: D  
Doc Type: ZSP

**REVISION HISTORY:**

<u>Revision</u>	<u>Description Change:</u>	<u>Changes Made By:</u>
D	Initial release embecta Form. Transitioned from existing Diabetes Care document 149OTH-0004-31 Ver. C. Removed ClickStar and added TouStar to the list of Sanofi pen injectors. Replaced BD references throughout document with embecta. Added embecta brands to list of embecta pen needles. Updated contact information for both Sanofi and embecta.	Jan Adlawan

CONFIDENTIAL - THE SUBJECT MATTER CONTAINED IN THIS DOCUMENT IS SOLELY FOR THE USE OF embecta  
The reader is responsible for ensuring that they use only the current revision of this document for compliance purposes